


ECHO HOOD® APERTURE-2

BALLOON 2 OPENINGS

45



The ECHO HOOD® APERTURE-2 balloon for echo-endoscope fits on the ultrasound probe to make contact with the gastrointestinal wall and thus provide a good ultrasound image.

ADVANTAGES

- Can withstand handling.
- Adequate thickness for the passage of ultrasound waves.
- Product compatible for use with the main endoscopes on the market.

PRECAUTIONS FOR USE

DO NOT USE IN PATIENTS ALLERGIC TO LATEX

AVAILABLE REFERENCE(S)

Reference	Compatible endoscopes	Length of the balloon (mm)	Size of 1 st opening (mm)	Size of 2 nd opening (mm)
AS1070000	ECHO-DUODENOSCOPE Olympus	16	5.4	2.7
AS2070000	ECHO-DUODENOSCOPE Pentax	19	6.5	5
AS3710000	ECHO-DUODENOSCOPE Pentax	22	1.8	4.2
AS2660000	ECHO-ENDOSCOPE Fujinon radial	19	5.8	5.8
AS3610000	ECHO-ENDOSCOPE Fujinon radial	19	6	6
AS2230000	ECHO-ENDOSCOPE series 160 Olympus	17	5.4	2.7
AS1460000	ECHOCOLOSCOPE Olympus	17	6	8.8
AS3280000	ECHO-BRONCHOSCOPE Olympus	12	1.3	3.3
AS3940000	ECHO-BRONCHOSCOPE Pentax	13.5	3.1	1.15

ABOUT THIS MEDICAL DEVICE

1 st release date	1998
Single use	Yes
Class	Is
Sterile	Yes
Sterilisation method	Ethylene oxide sterilisation
Contains latex	Yes
Lifetime	2 years
Certification	CE
Packaging	Box of 20 units in individual sterile wrapping
Storage	In its original box, protected from sunlight, in a cool and dry place

PRINCE MEDICAL SAS
 ZA La Sente Du Moulin
 64 rue Gutenberg - 60530 Ercuis - France
www.prince-medical.com
 Tel: +33 (0)3 44 26 54 76
prince-medical@omerin.com

The information provided on this data sheet is indicative and subject to change without notice. PRINCE MEDICAL SAS shall not be held responsible for any incidents resulting from inappropriate use, failure to comply with the rules and regulations in force, or failure to follow the instructions for use provided. Medical devices sold by PRINCE MEDICAL SAS should only be used by a duly qualified practitioner or authorised person.

® Registered trademark of PRINCE MEDICAL SAS. Drawings and photos are not contractually binding. Reproduction is prohibited without the prior consent of PRINCE MEDICAL SAS.

PRINCE MEDICAL
 DEVICES